

QIASymphony SP Protocol Sheet

Complex200_OBL_V4_DSP protocol

General information

For in vitro diagnostic use.

Kit	QIASymphony® DSP Virus/Pathogen Mini Kit
Sample material	Respiratory and urogenital samples
Protocol name	Complex200_OBL_V4_DSP
Default Assay Control Set	ACS_Complex200_OBL_V4_DSP
Editable	Eluate volume: 60 μ l, 85 μ l, 110 μ l
Required software version	Version 4.0

“Sample” drawer

Sample type	Respiratory samples (BAL, dried swabs, transport media, aspirates, sputum) and urogenital samples (urine, transport media)
Sample volume	Depends on type of sample tube used; for more information see www.qiagen.com/goto/dsphandbooks
Primary sample tubes	See www.qiagen.com/goto/dsphandbooks for more information
Secondary sample tubes	See www.qiagen.com/goto/dsphandbooks for more information
Inserts	Depends on type of sample tube used; for more information see www.qiagen.com/goto/dsphandbooks
Other	Carrier RNA–Buffer AVE mix required; use of internal control is optional

March 2012



Sample & Assay Technologies

“Reagents and Consumables” drawer

Position A1 and/or A2	Reagent cartridge (RC)
Position B1	n/a
Tip rack holder 1–17	Disposable filter-tips, 200 μ l
Tip rack holder 1–17	Disposable filter-tips, 1500 μ l
Unit box holder 1–4	Unit boxes containing sample prep cartridges
Unit box holder 1–4	Unit boxes containing 8-Rod Covers

n/a = not applicable.

“Waste” drawer

Unit box holder 1–4	Empty unit boxes
Waste bag holder	Waste bag
Liquid waste bottle holder	Liquid waste bottle

“Eluate” drawer

Elution rack (we recommend using slot 1, cooling position)	See www.qiagen.com/goto/dsphandbooks for more information
---	---

Required plasticware

	One batch, 24 samples*	Two batches, 48 samples*	Three batches, 72 samples*	Four batches, 96 samples*
Disposable filter-tips, 200 μ l†‡	96	96	128	128
Disposable filter-tips, 1500 μ l†‡	128	192	224	288
Sample prep cartridges§	18	36	54	72
8-Rod Covers¶	3	6	9	12

* Performing more than one inventory scan requires additional disposable filter tips. Use of less than 24 samples per batch decreases the number of disposable tips required per run.

† There are 32 filter-tips/tip rack.

‡ Number of required filter-tips includes filter-tips for 1 inventory scan per reagent cartridge.

§ There are 28 sample prep cartridges/unit box.

¶ There are twelve 8-Rod Covers/unit box.

Note: Numbers of filter-tips given may differ from the numbers displayed in the touchscreen depending on settings, for example, number of internal controls used per batch.

Selected elution volume

Selected elution volume (μ l)*	Initial elution volume (μ l)†
60	90
85	115
110	140

* The elution volume selected in the touchscreen. This is the minimum accessible volume of eluate in the final elution tube.

† The initial volume of elution solution required to ensure that the actual volume of eluate is the same as the selected volume.

Preparation of internal control–carrier RNA (CARRIER)–Buffer AVE (AVE) mixture

Selected elution volume (μ l)	Volume stock carrier RNA (CARRIER) (μ l)	Volume internal control (μ l)*	Volume Buffer AVE (AVE) (μ l)	Final volume per sample (μ l)
60	3	9	108	120
85	3	11.5	105.5	120
110	3	14	103	120

* The calculation of the amount of internal control is based on the initial elution volumes. Additional void volume depends on the type of sample tube used; see www.qiagen.com/goto/dsp/handbooks for more information.

Note: The values displayed in the table are for preparation of internal control–carrier RNA (CARRIER) mixture for a downstream assay that requires 0.1 μ l internal control/ μ l eluate.

Off-board lysis

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, consult the appropriate material safety data sheets (MSDSs), available from the product supplier.

The QIASymphony Complex protocols consist of 4 steps: lyse, bind, wash, elute. For some samples it is useful to perform lysis manually, for example, for inactivation of pathogens in a biosafety cabinet. The Complex200_OBL_V4_DSP protocol enables manual lysis to be performed in a similar way as for the Complex200_V6_DSP protocol. Pretreated samples are transferred to the QIASymphony SP and processed with the Complex200_OBL_V4_DSP.

Note: The Complex200_OBL_V4 protocol requires Buffer ACL and Buffer ATL (ATL). Buffer ACL (cat. no. 939017) and Buffer ATL (ATL) (cat. no. 939016) are not part of the QIASymphony DSP Virus/Pathogen Mini Kit and must be ordered separately.

Manual lysis

1. **Pipet 20 μ l proteinase K, 100 μ l Buffer ATL (ATL), 120 μ l Carrier RNA Internal Control Mixture, and 190 μ l Buffer ACL into a 2 ml Sarstedt tube (cat. no. 72.693 or 72.694).**

Note: When more than one sample will be processed using manual lysis, a stock solution of this solution can be prepared. Simply multiply the volumes required for one sample by the total number of samples to be processed, and include additional volume to the equivalent of 2 extra samples. Invert the tube several times to mix, transfer 430 μ l to a 2 ml Sarstedt tube for each sample, and then continue for each sample with step 4.

2. Close the lid and mix by inverting the tube 5 times.
3. Briefly centrifuge the tube to remove droplets from inside the lid.
4. Add 200 μ l sample to the tube, close the lid, and mix by pulse-vortexing for 10 seconds.
5. Incubate the tube at 68°C for 15 minutes (\pm 1 minute).
6. Briefly centrifuge the tube to remove droplets from inside the lid.
7. Place the inserts for the appropriate sample tubes into a tube carrier and load the sample tubes (without lids).

Preparation of sample material

Urine

Urine can be processed without further pretreatment. The system is optimized for pure urine samples that do not contain preservatives. To increase sensitivity for bacterial pathogens, samples can be centrifuged. After discarding the supernatant, the pellet can be resuspended in at least 200 μ l Buffer ATL (ATL) (cat. no. 939016). Use 200 μ l of the pre-treated material as sample for preparation of the off-board lysis.

Isolation of genomic DNA from Gram-positive bacteria

DNA purification can be improved for some Gram-positive bacteria by enzymatic pretreatment before transferring the sample to the QIA Symphony SP and starting the Complex200_OBL_V4_DSP protocol.

1. Pellet bacteria by centrifugation at 5000 x g for 10 minutes.
2. Suspend the bacterial pellet in 200 μ l of the appropriate enzyme solution (20 mg/ml lysozyme or 200 μ g/ml lysostaphin in 20 mM Tris-HCl, pH 8.0; 2 mM EDTA; 1.2% Triton X-100).
3. Incubate at 37°C for at least 30 minutes (\pm 2 minutes).
4. Briefly centrifuge the tube to remove drops from the inside of the lid.
5. Use 200 μ l of the pre-treated material as sample for preparation of the off-board lysis.

Viscous or mucous samples

Some samples (e.g., sputum, respiratory aspirates) may be viscous and require liquefaction to enable pipetting. Low-viscosity samples do not require additional preparation. Medium- to high-viscosity samples should be prepared as follows:

1. **Dilute the sample 1:1 with Sputasol*† (Oxoid, cat. no. SR0233) or 0.3% (w/v) DTT.**

Note: The 0.3 % DTT solution can be made in advance, and stored at –20°C in appropriate aliquots. Thawed aliquots should be discarded after use.

2. **Incubate at 37°C until the sample viscosity is suitable for pipetting.**
3. **Use 200 µl of the pre-treated material as sample for preparation of the off-board lysis.**

Dried body fluid and secretion swabs

1. **Submerge the dried swab tip in 450 µl Buffer ATL (ATL) (cat. no. 939016), and incubate at 56°C for 15 minutes (± 1 minute), with continuous mixing. If mixing is not possible, vortex before and after incubation for at least 10 seconds.**
2. **Remove the swab and squeeze out all the liquid by pressing the swab against the inside of the tube.**
3. **Use 200 µl of the pre-treated material as sample for preparation of the off-board lysis.**

Note: This protocol is optimized for cotton or polyethylene swabs. When using other swabs, it may be necessary to adjust the volume of Buffer ATL (ATL) to ensure that at least 200 µl is available as sample material.

Respiratory or urogenital swabs

Storage media for respiratory or urogenital swabs can be used without pretreatment. If the swab has not been removed, press the swab against the side of the tube to express the liquid. Any excess mucus in the specimen should be removed at this time by collecting it on the swab. Any residual liquid from the mucus and the swab should then be expressed by pressing the swab against the side of the tube. Finally, the swab and the mucus should be removed and discarded. If samples are viscous, perform a liquefaction step (see “Viscous or mucus samples” above) before transferring the sample to the QIA Symphony SP. If there is not sufficient starting material, pipet Buffer ATL (ATL) into the transport medium to adjust the required minimum starting volume and vortex the sample for 15–30 seconds in the tube (if the transport medium contains the swab perform this step before removing the swab). Use 200 µl of the material as sample for preparation of the off-board lysis.

* Sputasol (Oxoid, cat. no. SR0233, www.oxoid.com) or dithiothreitol (DTT).

† This is not a complete list of suppliers.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN handbooks can be requested from QIAGEN Technical Service or your local QIAGEN distributor. Selected handbooks can be downloaded from www.qiagen.com/literature. Material safety data sheets (MSDS) for any QIAGEN product can be downloaded from www.qiagen.com/Support/MSDS.aspx.

Trademarks: QIAGEN®, QIASymphony® (QIAGEN Group). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.
© 2012 QIAGEN, all rights reserved.

www.qiagen.com

Australia ■ 1-800-243-800

Austria ■ 0800/281010

Belgium ■ 0800-79612

Canada ■ 800-572-9613

China ■ 021-51345678

Denmark ■ 80-885945

Finland ■ 0800-914416

France ■ 01-60-920-930

Germany ■ 02103-29-12000

Hong Kong ■ 800 933 965

Ireland ■ 1800 555 049

Italy ■ 800 787980

Japan ■ 03-5547-0811

Korea (South) ■ 1544 7145

Luxembourg ■ 8002 2076

The Netherlands ■ 0800 0229592

Norway ■ 800-18859

Singapore ■ 65-67775366

Spain ■ 91-630-7050

Sweden ■ 020-790282

Switzerland ■ 055-254-22-11

UK ■ 01293-422-911

USA ■ 800-426-8157



Sample & Assay Technologies