

# QIAGEN Supplementary Protocol

---

## Singleplex and duplex analysis using the QuantiTect<sup>®</sup> Virus Kit and a 25 $\mu$ l reaction volume

This protocol shows how to use the QuantiTect Virus Kit to perform real-time PCR in a singleplex or duplex format with a 25  $\mu$ l reaction volume. The protocol is intended for use with TaqMan<sup>®</sup> probes and the following real-time cyclers from Applied Biosystems: ABI PRISM<sup>®</sup> 7000, Applied Biosystems<sup>®</sup> 7300, ABI PRISM 7700, ABI PRISM 7900HT, Applied Biosystems 7900HT Fast, and StepOnePlus<sup>™</sup>. For details, see “ROX passive reference dye” in the *QuantiTect Virus Handbook*.

**IMPORTANT:** Please consult the “Safety Information” and “Important Notes” sections in the *QuantiTect Virus Handbook* before beginning this procedure. For safety information on the additional chemicals mentioned in this protocol, please consult the appropriate material safety data sheets (MSDSs), available from the product supplier.

### Equipment and reagents

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, consult the appropriate material safety data sheets (MSDSs), available from the product supplier.

- QuantiTect Virus Kit (cat. no. 211011, 211013, or 211015)
- Primers and probes: These should be purchased from an established oligonucleotide manufacturer. Primers should be of standard quality, and probes should be HPLC purified. Lyophilized primers and probes should be dissolved in TE buffer to provide a stock solution of 100  $\mu$ M; concentration should be checked by spectrophotometry (for details, see Appendix A in the *QuantiTect Virus Handbook*). Primer and probe stock solutions should be stored in aliquots at  $-20^{\circ}\text{C}$ . Probe stock solutions should be protected from exposure to light.
- Nuclease-free (RNase/DNase free) consumables: Special care should be taken to avoid nuclease contamination of all reagents and consumables used to set up PCR for sensitive detection of viral nucleic acids. See Appendix C in the *QuantiTect Virus Handbook* for details about avoiding nucleases during PCR setup.
- Cooling device or ice
- Real-time thermal cycler
- PCR tubes or plates (use thin-walled PCR tubes or plates recommended by the manufacturer of your thermal cycler)



- Optional: Trizma® base and EDTA for preparing TE buffer for storing primers and probes (see Appendix A in the *QuantiTect Virus Handbook*). Use RNase/DNase-free water and plastic consumables to prepare TE buffer.

## Important points before starting

- **Always start with the cycling conditions specified in this protocol.** The cycling is optimized for PCR products between 60 and 150 bp. For PCR products > 150 bp, different cycling conditions may improve results. For details, see Appendix A in the *QuantiTect Virus Handbook*.
- **Use the primer concentrations specified in this protocol.**
- We strongly recommend testing the performance of new primer–probe sets in individual assays before combining them in a multiplex assay.
- Read “Guidelines for effective multiplex assays” in the *QuantiTect Virus Handbook*. Check whether your real-time cyler is compatible with the chosen combination of reporter dyes.
- See the user manual for the real-time thermal cyler for correct instrument setup for multiplex analysis (e.g., setting up detection of multiple dyes in the same well). Make sure to activate the detector for each reporter dye used. Depending on the instrument, it may also be necessary to perform a calibration procedure for each of the reporter dyes before they are used for the first time.
- If using an already established real-time, duplex assay, use the previously established primer and probe concentrations in combination with the cycling conditions specified in this protocol. It is not necessary to determine primer-limiting concentrations again.
- **After reverse transcription, the PCR must start with an initial incubation step of 5 min at 95°C to activate HotStarTaq® Plus DNA Polymerase.**
- **Optimal analysis settings are a prerequisite for accurate quantification data.** For data analysis, you should always readjust the analysis settings (i.e., baseline settings and threshold values) for analysis of every reporter dye channel in every run.

## Things to do before starting

- For ease of use, we recommend preparing a 20x primer–probe mix for each of your targets containing target-specific primers and probe. See Appendix D in the *QuantiTect Virus Handbook*. Alternatively, it may be preferable to prepare the reaction mix with separate primer and probe solutions. If you commonly set up reactions this way, see Appendix E in the *QuantiTect Virus Handbook*.

## Procedure

1. **Thaw QuantiTect Virus Master Mix, primer and probe solutions, RNase-free water, template nucleic acids (isolated viral nucleic acids), optional standards, and references. Mix the individual solutions.**

Standards should be diluted in QuantiTect Nucleic Acid Dilution Buffer at an appropriate concentration to enable use of 2.5–5  $\mu$ l per reaction.

2. **Prepare a reaction mix for the required number of reactions according to Table 1. It is recommended to prepare a volume of reaction mix 10% greater than that required for the total number of reactions to be performed.**

Typically, reaction setup can be done at room temperature (15–25°C). However, it is recommended to keep the reagents, samples, and controls on ice or in a cooling device.

**Note:** For RT-PCR, QuantiTect Virus RT Mix should be taken from –15 to –30°C immediately before use, always kept on ice, and returned to –15 to –30°C immediately after use.

3. **Mix the reaction mix thoroughly, and dispense appropriate volumes into PCR tubes or the wells of a PCR plate.**
4. **Add template nucleic acids to the individual PCR tubes or wells and mix thoroughly.**

**Note:** Ensure that the reaction mix and template are thoroughly mixed.
5. **Program the real-time cycler according to Table 2.**

**Table 1. Reaction mix for singleplex or duplex analysis**

<b>Component</b>	<b>Volume per 25 <math>\mu</math>l reaction*</b>	<b>Final concentration</b>
<b>Reaction mix</b>		
QuantiTect Virus Master Mix, 5x	5 $\mu$ l	1x
20x primer–probe mix 1 <sup>†</sup>	1.25 $\mu$ l	0.4 $\mu$ M forward primer 1 <sup>‡</sup> 0.4 $\mu$ M reverse primer 1 <sup>‡</sup> 0.2 $\mu$ M probe 1 <sup>§</sup>
20x primer–probe mix 2 <sup>†</sup>	1.25 $\mu$ l	0.4 $\mu$ M forward primer 2 <sup>‡</sup> 0.4 $\mu$ M reverse primer 2 <sup>‡</sup> 0.2 $\mu$ M probe 2 <sup>§</sup>
<b>For RT-PCR only:</b> QuantiTect Virus RT Mix, 100x	0.25 $\mu$ l	1x
RNase-free water	Variable	–
<b>Template DNA or RNA</b> (added at step 4)	Variable	Maximum up to 50% of final reaction volume
<b>Total volume per reaction</b>	<b>25 <math>\mu</math>l*</b>	–

\* If your real-time cycler requires a final reaction volume other than 25  $\mu$ l, adjust the amount of master mix and all other reaction components accordingly.

<sup>†</sup> For ease of use, we recommend preparing a 20x primer–probe mix for each of your targets containing target-specific primers and probe. See Appendix D in the *QuantiTect Virus Handbook*.

<sup>‡</sup> A final primer concentration of 0.4  $\mu$ M is optimal. Before adapting primer concentration, verify the concentration of your primer solutions.

<sup>§</sup> A final probe concentration of 0.2  $\mu$ M gives satisfactory results in most cases. Depending on the synthesis quality and purification method used, the optimal concentration may be between 0.1  $\mu$ M and 0.4  $\mu$ M.

**Table 2. Cycling conditions for singleplex or duplex analysis**

Step	Time	Temperature	Additional comments
<b>For RT-PCR only: Reverse transcription</b>	<b>20 min</b>	<b>50°C</b>	RNA is reverse transcribed into cDNA. Omit this step if you are analyzing DNA targets.
<b>Initial PCR activation step</b>	<b>5 min</b>	<b>95°C</b>	HotStarTaq <i>Plus</i> DNA Polymerase is activated by this heating step.
<b>2-step cycling:</b>		<b>Important: Optimal performance is only assured using these cycling conditions.</b>	
Denaturation	<b>15 s</b>	<b>95°C</b>	
Annealing/extension	<b>45 s</b>	<b>60°C</b>	Combined annealing/extension step with fluorescence data collection, optimized for PCR products up to 150 bp. For PCR products >150 bp, different cycling conditions may improve results in some cases. For details, see Appendix A in the <i>QuantiTect Virus Handbook</i> .
Number of cycles	40–50		The number of cycles depends on the amount of template DNA or RNA.

**6. Place the PCR tubes or plate in the real-time cycler and start the PCR cycling program.**

**7. Perform data analysis.**

Before performing data analysis, specify the analysis settings. For each probe, select the analysis settings (i.e., baseline settings and threshold values). Note that optimal analysis settings are a prerequisite for accurate quantification data.

QIAGEN handbooks can be requested from QIAGEN Technical Service or your local QIAGEN distributor. Selected handbooks can be downloaded from [www.qiagen.com/literature](http://www.qiagen.com/literature).

Material safety data sheets (MSDS) for any QIAGEN product can be downloaded from [www.qiagen.com/Support/MSDS.aspx](http://www.qiagen.com/Support/MSDS.aspx).

QuantiTect Virus Kits are intended for research use. No claim or representation is intended to provide information for the diagnosis, prevention, or treatment of a disease.

Trademarks: QIAGEN®, HotStarTaq®, QuantiTect® (QIAGEN Group); ABI PRISM®, Applied Biosystems®, StepOnePlus™ (Applied Biosystems Corporation or its subsidiaries); TaqMan® (Roche Group); Trizma® (Sigma-Aldrich Co.). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

#### NOTICE TO PURCHASER: LIMITED LICENSE

A license to perform the 5' nuclease process for research requires the use of a Licensed 5' Nuclease Kit (containing Licensed Probe), or the combination of an Authorized 5' Nuclease Core Kit plus Licensed Probe, or license rights that may be purchased from Applied Biosystems. This product (QuantiTect Virus Kit) is an Authorized 5' Nuclease Core Kit without Licensed Probe. Its purchase price includes a limited, non-transferable immunity from suit under U.S. Patents Nos. 5,210,015, 5,487,972, 5,476,774, and 5,219,727, and corresponding patent claims outside the United States, owned by Roche Molecular Systems, Inc. or F. Hoffmann-La Roche Ltd (Roche), for using only this amount of the product in the practice of the 5' nuclease process solely for the purchaser's own internal research when used in conjunction with Licensed Probe. This product is also an Authorized 5' Nuclease Core Kit for use with service sublicenses available from Applied Biosystems. This product conveys no rights under U.S. Patents Nos. 5,804,375, 6,214,979, 5,538,848, 5,723,591, 5,876,930, 6,030,787, or 6,258,569, or corresponding patents outside the United States, expressly, by implication or by estoppel. No right under any other patent claims (such as apparatus or system claims in U.S. Patent No. 6,814,934) and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is hereby granted expressly, by implication or by estoppel. This product is for research use only. Diagnostic uses require a separate license from Roche. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

#### NOTICE TO PURCHASER: LIMITED LICENSE

The purchase price of this product (QuantiTect Virus Kit) includes a limited, non-transferable license under U.S. Patents Nos. 5,407,800, 5,322,770, 5,310,652 and corresponding patent claims outside the United States, owned by Roche Molecular Systems, Inc. or F. Hoffmann-La Roche Ltd (Roche), to use only this amount of product solely for the purchaser's own internal research. No right under any other patent claims (such as apparatus or system claims) and no right to use this product for any other purpose or for commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is hereby granted expressly, by implication or by estoppel. This product is for research use only. Diagnostic uses require a separate license from Roche. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

Certain specific embodiments of the process of multiplex PCR may be covered by patents of third parties in certain countries and may require a license.

PCR88 Aug-08 © 2008 QIAGEN, all rights reserved.

[www.qiagen.com](http://www.qiagen.com)

**Australia** = 1-800-243-800

**Austria** = 0800/281010

**Belgium** = 0800-79612

**Canada** = 800-572-9613

**China** = 0086 21 3865 3865

**Denmark** = 80-885945

**Finland** = 0800-914416

**France** = 01-60-920-930

**Germany** = 02103-29-12000

**Hong Kong** = 800 933 965

**Ireland** = 1800 555 049

**Italy** = 800 787980

**Japan** = 03-5547-0811

**Korea (South)** = 1544 7145

**Luxembourg** = 8002 2076

**The Netherlands** = 0800 0229592

**Norway** = 800-18859

**Singapore** = 65-67775366

**Spain** = 91-630-7050

**Sweden** = 020-790282

**Switzerland** = 055-254-22-11

**UK** = 01293-422-911

**USA** = 800-426-8157



Sample & Assay Technologies