



Product Sheet

November 2022

DNeasy[®] PowerSoil[®] Pro QIAcube[®] Kit

For use on QIAcube Connect

| DNeasy PowerSoil Pro QIAcube Kit | (240) |
|----------------------------------|--------------------|
| Catalog no. | 47126 |
| Mini Spin Columns | 240 |
| Rotor Adapters | 240 |
| Elution Tubes | 240 |
| PowerBead Pro Tubes | 240 |
| Solution CD1* | 200 mL |
| Solution CD2 | 2 x 60 mL |
| Solution CD3* | 175 mL + 2 x 35 mL |
| Solution C5† | 2 x 175 mL |
| Solution C6 | 2 x 66 mL |
| Solution EA† | 175 mL |

* **CAUTION:** Contain chaotropic salts. Take appropriate laboratory safety measures and wear gloves when handling. Not compatible with disinfectants containing bleach. See page 2 for Safety Information.

† Flammable.

Shipping and Storage

Solution CD2 should be stored at 2–8°C upon arrival. All other components and reagents of the DNeasy PowerSoil Pro Kit can be stored at room temperature (15–25°C) until the expiration date printed on the box label.

Intended Use


All DNeasy products are intended for molecular biology applications. These products are not intended for the diagnosis, prevention, or treatment of a disease.


The QIAcube Connect is designed to perform fully automated purification of nucleic acids and proteins in molecular biology applications. The system is intended for use by professional users trained in molecular biological techniques and the operation of the QIAcube Connect.

All due care and attention should be exercised in the handling of the products. We recommend all users of QIAGEN® products to adhere to the NIH guidelines that have been developed for recombinant DNA experiments, or to other applicable guidelines.

Safety Information

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available online in convenient and compact PDF format at www.qiagen.com/Support/MSDS.aspx where you can find, view, and print the SDS for each QIAGEN kit and kit components.

| | |
|---------------------------------------------------------------------------------------------------------|--------------------------------------------|
| <p>WARNING</p>  | Solution EA and Solution C5 are flammable. |
|---------------------------------------------------------------------------------------------------------|--------------------------------------------|

| | |
|---------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------|
| <p>CAUTION</p>  | DO NOT add bleach or acidic solutions directly to the sample preparation waste. |
|---------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------|

Solution CD1 and Solution CD3 contain chaotropic salts, which can form highly reactive compounds when combined with bleach. If liquid containing these buffers is spilt, clean with a suitable laboratory detergent and water. If the spilt liquid contains potentially infectious agents, clean the affected area first with laboratory detergent and water, and then with 1% (v/v) sodium hypochlorite.

Document Revision History

| Date | Changes |
|---------|------------------|
| 11/2022 | Initial revision |

Trademarks: QIAGEN®, Sample to Insight®, QIAcube®, DNeasy®, PowerSoil® (QIAGEN Group). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

11/2022 HB-3149-001 © 2022 QIAGEN, all rights reserved.